



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: May 8, 2012 INITIALS: Dec OWNER: WASHINGTON STATE PATROL

DISTRICT: 3 ADDRESS: SR82 MP97

LOCATION: Grandview 28 North

COUNTY: Benton

MAKE: GSE

TYPE: Trk Scale

ACCURACY CLASS: III L

MODEL: 651

ID. NO: 202189

CAPACITY: 200000 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 20 lbs.

REG. ELEMENT: ELECTRONIC

PLATFORM 10X22

North

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
P1-Sec 1	20000	20020	+20	P2-Sec1	20000	20000	0
P1-Center	20000	20000	0	P2-Center	20000	20000	0
P1-Sec2	20000	20000	0	P2-Sec2	20000	20000	0

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L3184-1 and 200446-0-L3184-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.


Signed by:

David Cromer

Date:

June 14, 2012

Print name: David Cromer



State of Washington

County of Thurston

Signed and sworn to (or affirmed before me on June 14, 2012)

Patricia A. Geppert

(Signature of notary public)

PATRICIA A. GEPPERT

Printed Name of Notary

My appointment expires Sept. 4, 2015